



RECORD SHEET

Probe thermometer check log

Business: _____ Period: _____

0°C in iced water.**100°C in boiling water.**

Accepted check ranges: -1°C to +1°C and 99°C to 101°C.* Record the actual reading, not just a tick. Check at your planned frequency and after damage.

Check frequency: _____ Probe make / ID: _____

Date	Probe ID	Iced water °C	Boiling water °C	Pass / fail, action & recheck	Initials

Outside range? Stop using the probe for safety checks. Use a verified replacement and arrange calibration or replacement. Record the outcome.

*Boiling point falls with altitude. Use the expected local boiling point and manufacturer guidance. Take care with steam and hot water.